



REF 95702 Approximate priming volume 0.9 mL 1 UNIT



QUEST Medical, Inc.
An **Atrion** company

ALLEN, TEXAS 75002 USA
972 390-9800 / 800 627-0226
www.questmedical.com

EC REP

CE
0123

M. Devices Group
Marlborough House
Riding Street
Southport, PR8, 1EW, UK
Tel: +44 1704 544 944
Fax: +44 1704 544 050

©2010 Quest Medical, Inc. CheckMate, MultiPort,
and Q2 are trademarks of Quest Medical, Inc.

903842 REV E
APR 10

Instructions For Use

INTENDED USE:

For administration of intravenous fluids to a patient's vascular system utilizing needle-free components and I.V. manifolds for multiple simultaneous intravenous therapy via gravity, syringe, or infusion pump.

CONTRAINDICATIONS:

None known.

WARNINGS:

- 1) Intended for Single Patient Use only. DO NOT REUSE.
- 2) After use, this product may be a potential biohazard. Handle and discard in accordance with the accepted medical practice and applicable local, state, and federal laws and regulations.
- 3) For unattended medication deliver, only luer locking secondary sets and syringes should be used with injection sites.

CAUTIONS:

- 1) Before using, examine the package carefully before opening to confirm its integrity and that the expiration date has not passed. The device is supplied in a sterile package and is non-pyrogenic. DO NOT USE if package is damaged, opened, or the expiration date has passed. Do not resterilize.
- 2) Do not use needles or blunt cannulae with the needle-free injection sites.
- 3) Replace set per CDC guidelines or hospital protocol.
- 4) Flush injection sites in accordance with hospital policy after or between injections.
- 5) To prevent retrograde fluid flow after flushing the Q2® injection site with a syringe, twist syringe counterclockwise and continue depressing syringe plunger until syringe is disengaged from site.

POSSIBLE PATIENT COMPLICATIONS:

The potential exists for serious complications including the following:

- 1) Air embolism