ARISURE®
CLOSED SYSTEM TRANSFER DEVICE

ARISURE® BASIC ADMINISTRATION

Always follow the instructions detailed in the Arisure Directions For Use.

COMPONENTS*:

- YM104 – Arisure Closed Male Luer (CML)
- YM004 – Arisure Neutral Valve (NV) (or another compatible needle free connector, e.g., BD SmartSite** or Baxter One-link***)
- Medication tubing set
- Medication syringe

All hazardous drugs, as determined by NIOSH and your facility, will come from your pharmacy with the CML pre-attached, per facility guidelines (exceptions may exist for specific applications).

STEP 1: Add a NV or compatible needle free connector to patient's vascular access device or primary tubing access point (Y-site).
NV can be stacked onto incompatible bonded needle free connector if needed.

STEP 2: Obtain hazardous drug prepared with the CML from pharmacy.

STEP 3: Connect the CML to the NV, open clamps and infuse per protocol.
Reminder: attach NV to the patient first before attaching medication to prevent accidental spills.

STEP 4: Upon completion of administration flush per facility protocol and dispose of all exposed components in hazardous waste disposal bin per facility protocol.

*Per USP 800: Don PPE before handling any drugs deemed hazardous by NIOSH. Follow aseptic technique and abide by safe handling of hazardous drug standards.
**SmartSite is a trademark of Becton, Dickinson and Company, its subsidiaries or affiliates. Yukon Medical Inc. is not affiliated with Becton, Dickinson and Company.
***ONE-LINK AND CLEARLINK are trademarks of Baxter International Inc, its subsidiaries or affiliates. Yukon Medical Inc. is not affiliated with Baxter International Inc.

YUKON MEDICAL
The information in this sheet is not intended to be exhaustive. The information is liable to modification from time to time in light of experience and our commitment to continuous product development. These documents, in part or in whole, may not be copied or reproduced without prior written permission from Yukon Medical.